

MANUAL WHEELCHAIRS

Summary of Technical Tests

Product name: Easy Roller
ISO-code: 12 22 03

Client: Easy Roller AS
Sørlandsveien 50
8640 Hemnesberget
Norway

Manufacturer: Easy Roller AS
Sørlandsveien 50
8640 Hemnesberget
Norway



The wheelchair has been tested in accordance with EN 12183:2009 including ISO 7176-1, ISO 7176-3 ISO 7176-8 and EN 12183:1999 7.6.
It has also been tested in accordance with ISO 7176-5 and ISO 7176-7.

The tested wheelchair had a seat width of 45 cm and was equipped with the following elements (nomenclature as in ISO 6440): foldable leg support, fixed armrest and bar type handle.

During the tests adjustable parts were adjusted for driving according to the clients recommendations. For further information see "Identification List Part 1 and 2".

The tests was performed with a 100 kg test dummy in accordance with ISO 7176-11 with complementary weights up to 150 kg.

This report consist of 12 parts and 29 pages.

Test results: The wheelchair pass all applicable requirements given in the standards enumerated above. All results concern the tested products only.

Swedish Institute of Assistive Technology

2010-07-12

Issued by:

Approved by:



Svenne Magnusson, Test performer



Jennie Josefsson, Head of Testing Laboratory